

PROTOCOL: ELBOW ARTHROGRAM**COIL:** TRANSMIT-RECEIVE KNEE COIL WITH ARM ABOVE HEAD OR SMALL/MEDIUM 4+ CHANNEL FLEX COIL WRAPPED *AROUND* ELBOW. DO NOT LAY ELBOW ON**POSITIONING:** Supinated (palm up). Superman or arm at side. If arm at side, careful attention to avoid phase wrap.**CLINICAL INDICATIONS/** ELBOW PAIN, BICEPS TEAR, TENNIS ELBOW, EPIDONDYLITIS, INSTABILITY, INTERNAL DERANGEMENT, ARTHRITIS**PLOTTING:** Axial FIRST off of coronal and sagittal localizers. True axial to humerus Coronal = linear cortex between coronoid fossa and olecranon fossa Sagittal =**COVERAGE:** Above medial and lateral epicondyles. Above olecranon. Below radial tuberosity (must see entire biceps insertion).**TIPS:** Arm as straight as possible. If unable to straighten arm (near 90 degree flexion), repeat coronal & axial for both humerus and forearm

SCAN ORDER	PLANE	IMAGE CONTRAST/WEIGHTING	MODE/SEQ TYPE	Averages (NEX)	BW (Hz/pixel)	ETL	TR RANGE	TE RANGE	TI	FLIP ANGLE	SLICE/GAP (mm)	FOV (cm)	Resolution	Phase Axis	Send to PACS	Fat sat
1	AX	T1FS	2D/FSE	1	150-180	2-3	450-650	8-10		>130	3.0/0	10-13	256 x 256	RL	FULL SERIES	WEAK
2	AX	PD	2D/FSE	2	150-180	8-12	3500-6000	38-45		>130	3.0/0	10-13	512 X 320-384	RL	FULL SERIES	NONE
3	COR	T2FS	2D/FSE	1	150-180	8-12	3500-6000	65-75		>130	3.0/0	10-13	256 X 256	SI	FULL SERIES	WEAK
4	COR	PD	2D/FSE	2	150-180	8-12	3500-6000	38-45		>130	3.0/0	10-13	512 x 320-384	SI	FULL SERIES	NONE
5	COR	T1	2D/FSE	1	150-180	2-3	450-650	8-10		>130	3.0/0	10-13	256 x 192	SI	FULL SERIES	NONE
6	SAG	PDFS	2D/FSE	1	150-180	8-12	3500-6000	38-45		>130	3.0/0	10-13	320 X 288	SI	FULL SERIES	WEAK