

**PROTOCOL: MR PELVIS WWO-UTERUS/CERVIX/OVARIES (NOT TO BE USED FOR UAE PRE-EMBO)**

**CLINICAL INDICATIONS/ HISTORY:** Uterine Mass/abnormality, uterine cancer, cervical mass, cervical cancer, ovarian mass/cancer. Nonspecific clinical questions in males other than prostate. This protocol is used for all female pelvis indications including those below, except for uterine fibroid PRE-EMBOLIZATION exams. That is covered in a separate protocol. This protocol may also be used in male patients for diagnostic questions OTHER THAN prostate, which is a specific protocol.

SCAN ORDER	PLANE	IMAGE CONTRAST / WEIGHTING	MODE	PULSE SEQ	COVERAGE	TR RANGE	TE RANGE	FLIP ANGLE	THICKNESS / GAP (mm)	FOV (cm)	Max Pixel (mm) Fr x Ph	PHASE AXIS	SEND TO PACS	Max scan time (target)
1	COR	T2	2D	HASTE/SSFSE	ENTIRE PELVIS	>1000	80-100	>130	5/1.5	36-40	.9X1.5	IS	FULL SERIES	2:00
2	AX	T2	2D	FSE/ PROPELLER/ BLADE	TRUE PELVIS THRU VULVA	>3000	120	>130	4-5/1.5	24-28	.9X1.0	RL	FULL SERIES	4:00
3	SAG	T2	2D	FSE/ PROPELLER/ BLADE	AREA BETWEEN ACETABULA	>3000	120	>130	4-5/1.5	24-28	.9X1.0	SI	FULL SERIES	4:00
4	COR	T2 FS	2D	FSE/ PROPELLER/ BLADE	COCCYX THRU SYMPHYSIS	>3000	120	>130	4-5/1.5	24-28	.9X1.0	RL	FULL SERIES	4:00
5	AX	T1	2D	FSE	ENTIRE PELVIS	<790	MIN	>130	5/1.5	32-36	.9X1.5	RL	FULL SERIES	3:00
6	AX	T1	3D	VIBE DIXON/ LAVA FLEX	ENTIRE PELVIS	MIN	MIN	8-15	3-4/OVLP	32-36	.9X1.5	PA	IN/OUT/WATER	<18s
7	AX	DWI B50/800	2D	EPI/FOCUS/ RESOLVE	ENTIRE PELVIS	>5000	MIN	180	5/1.5	32-36	2.0X2.75	AP	FULL SERIES+ADC	4:00
8	AX	T1 FS	3D	VIBE DIXON/ LAVA FLEX	ENTIRE PELVIS	MIN	MIN	8-15	3-4/OVLP	32-36	1.4X1.7	PA	IN/OUT/WATER	<18s
9-10-11+C	AX	T1 FS DYNAMIC	3D	VIBE DIXON/ LAVA FLEX- DYN 3 PHASE: 25/70/120s	ENTIRE PELVIS	MIN	MIN	8-15	3-4/OVLP	32-36	1.4X1.7	PA	WATER	<18s
12+C	COR	T1 FS	3D	VIBE DIXON/ LAVA FLEX	ENTIRE PELVIS	MIN	MIN	8-15	3-4/OVLP	32-36	1.4X1.7	RL	WATER	3:00
13+C	SAG	T1 FS	3D	VIBE DIXON/ LAVA FLEX	ENTIRE PELVIS	MIN	MIN	8-15	3-4/OVLP	24-28	1.4X1.7	SI	WATER	3:00