

PROTOCOL: MR PELVIS WO-PENILE TRAUMA *POSITION PENIS PARALLEL TO ABDOMINAL WALL, USING WASHCLOTHS AND TAPE TO SECURE**

CLINICAL INDICATIONS/ HISTORY: PENILE FX/INJURY

SCAN ORDER	PLANE	IMAGE CONTRAST / WEIGHTING	MODE	PULSE SEQ	COVERAGE	TR RANGE	TE RANGE	FLIP ANGLE	THICKNESS / GAP (mm)	FOV (cm)	Max Pixel (mm) Fr x Ph	PHASE AXIS	SEND TO PACS	Max scan time (target)
1	COR	T2	2D	HASTE/SSFSE	ENTIRE PELVIS	>1000	80-100	>130	5/1.5	36-40	.9X1.5	IS	FULL SERIES	3:00
2	SHORT AXIS	T2	2D	FSE/ PROPELLER/ BLADE	PENIS/URETHRA	>3000	120	>130	4/.4	18-20	.9X1.0	RL	FULL SERIES	4:00
3	SAG	T2	2D	FSE/ PROPELLER/ BLADE	PENIS/URETHRA	>3000	120	>130	4/.4	18-20	.9X1.0	SI	FULL SERIES	4:00
4	COR-LONG AXIS	T2	2D	FSE/ PROPELLER/ BLADE	PENIS/URETHRA	>3000	120	>130	4/.4	18-20	.9X1.0	RL	FULL SERIES	4:00
5	SHORT AXIS	T1	2D	FSE	PENIS/URETHRA	<790	MIN	>130	4/.4	18-20	.9X1.0	RL	FULL SERIES	3:00
6	AX	T1 FS	3D	VIBE DIXON/ LAVA FLEX	PENIS/URETHRA	MIN	MIN	8-15	3-4/OVLP	24-32	1.4X1.7	PA	IN/OUT/WATER	<18s