

PROTOCOL: MR PELVIS WWO-URETHRA/PENIS *POSITION PENIS PARALLEL TO ABDOMINAL WALL, USING WASHCLOTHS AND TAPE TO SECURE*****

CLINICAL INDICATIONS/ HISTORY: URETHRAL DIVERTICULUM/PENILE MASS/INFECTION/CA

SCAN ORDER	PLANE	IMAGE CONTRAST / WEIGHTING	MODE	PULSE SEQ	COVERAGE	TR RANGE	TE RANGE	FLIP ANGLE	THICKNESS / GAP (mm)	FOV (cm)	Max Pixel (mm) Fr x Ph	PHASE AXIS	SEND TO PACS	Max scan time (target)
1	COR	T2	2D	HASTE/SSFSE	ENTIRE PELVIS	>1000	80-100	>130	5/1.5	36-40	.9X1.5	IS	FULL SERIES	3:00
2	SHORT AXIS	T2	2D	FSE/ PROPELLER/ BLADE	PENIS/ URETHRA	>3000	120	>130	4/.4	18-20	.9X1.0	RL	FULL SERIES	4:00
3	SAG	T2	2D	FSE/ PROPELLER/ BLADE	PENIS/ URETHRA	>3000	120	>130	4/.4	18-20	.9X1.0	SI	FULL SERIES	4:00
4	COR-LONG AXIS	T2	2D	FSE/ PROPELLER/ BLADE	PENIS/ URETHRA	>3000	120	>130	4/.4	18-20	.9X1.0	RL	FULL SERIES	4:00
5	SHORT AXIS	T1	2D	FSE	PENIS/ URETHRA	<790	MIN	>130	4/.4	18-20	.9X1.0	RL	FULL SERIES	3:00
6	AX	DWI B50/800	2D	EPI/FOCUS/ RESOLVE	PENIS/ URETHRA	>5000	MIN	180	5/1.5	24-32	2.0X2.75	AP	FULL SERIES+ADC	4:00
7	AX	T1 FS	3D	VIBE DIXON/ LAVA FLEX	PENIS/ URETHRA	MIN	MIN	8-15	3-4/OVLP	24-32	1.4X1.7	PA	IN/OUT/WATER	<18s
8-9-10+C	AX	T1 FS DYN	3D	VIBE DIXON/ LAVA FLEX	PENIS/ URETHRA	MIN	MIN	8-15	3-4/OVLP	24-32	1.4X1.7	PA	3PH 20-70-120	<18s
11+C	COR	T1 FS	3D	VIBE DIXON/ LAVA FLEX	PENIS/ URETHRA	MIN	MIN	8-15	3-4/OVLP	24-32	1.4X1.7	RL	WATER	<18s
12+C	SAG	T1 FS	3D	VIBE DIXON/ LAVA FLEX	PENIS/ URETHRA	MIN	MIN	8-15	3-4/OVLP	24-28	1.4X1.7	SI	WATER	<18s